

 **FARAPULSE™**
Pulsed Field Ablation System




FARAWAVE™
Pulsed Field Ablation Catheter



FARAWAVE™ Ordering Information

Reference Number	
41M401	31 mm (distal diameter)
41M402	35 mm (distal diameter)
Catheter Specifications	
Catheter type	Over-the-wire
Shaft diameter	12.8F
Usable length	115 cm
Number of electrodes	20
Compatibility	
Guidewire	0.035" diameter, 180 cm minimum length
Sheath	FARADRIVE™ (REF 21M402)
Accessories	
FARASTAR catheter connection cable	REF 41M404

FARAWAVE™ is sterile and single-use only






FARADrive™ Ordering Information

Reference Number	
21M402	Clear shaft
Sheath Specifications	
Usable Length	74 cm
Inner Diameter	13.0 F
Deflection	Unidirectional, 35 mm reach
Radio-Opaque Marker	Distal end
Dilator Specifications	
Total Length	97 cm
Usable Length	94 cm
Compatibility	
Guidewires	Maximum diameter 0.035"
Transseptal Needle	Abbott BRK™, 98 cm Merit Medical HeartSpan, 98 cm
Pulsed Field Ablation Catheter	FARAWAVE™ (REF 41M401, 41M402)

FARADrive™ Steerable Sheath and dilator are sterile and single-use only





FARASTAR™ Ordering Information

Reference Number	
61M401	FARASTAR Pulsed Field Ablation Generator
Shipping Information	
Weight	250 kg
Dimensions	L 120 x H 82 x W 198 cm
Technical Specifications	
Input Voltage	230V / 50Hz (EU)
Output Voltage	1.8 - 2.0 kV, user-adjustable Proprietary FARAPULSE biphasic, bipolar waveform
Weight	110 kg
Dimensions	L 79 x W 61 x H 117 cm (160 cm with monitor)
User Interface	Articulating touchscreen
Accessories	
61M404	FARASTAR Stimulation Module Cable
61M405	FARASTAR EGM Cable
61M406	FARASTAR Accessories Cable Set
61M407	FARASTAR Recording System Module
61M408	FARASTAR Stimulation Module Auxiliary Cable
61M409	FARASTAR Stimulation Module Male Cable
61M410	FARASTAR Stimulation Module Female Cable
61M411	FARASTAR Stimulation Module Y-Cable Long
61M412	FARASTAR Stimulation Module Y-Cable Short
61M413	FARASTAR Recording System Module Catheter Pin Cable
61M415	FARASTAR Recording System Module ECG Trunk Cable
61M416	FARASTAR Recording System Module ECG Output Module
61M417	FARASTAR Recording System Module EGM Input Module

CAUTION: The law restricts these devices to sale by or on the order of a physician. Indications, contraindications, warnings and instructions for use can be found in the product labelling supplied with each device. Products shown for INFORMATION purposes only and may not be approved or for sale in certain countries.

This material not intended for use in France.

© 2022 Boston Scientific Corporation or its affiliates. All rights reserved.

EP- 1161106-AB

**Boston
Scientific**

Advancing science for life™

www.bostonscientific.eu